



Tohnichi America Corp.

677 Academy Drive

Northbrook, IL 60062

TEL : 847 272 8480 FAX : 847 272 8714

CERTIFICATE OF CALIBRATION (SAMPLE)

Certificate # :	<u>041211-01</u>	Date First Used :	<u> / /</u>
Type of Tool :	<u>Torque Wrench</u>	Accuracy ± :	<u>3%</u>
Model :	<u>250CF</u>	Serial Number :	<u>XXXXXXA</u>
Units :	<u>kgf.cm</u>	Temperature F :	<u>71.2</u>
Date of Calibration :	<u>04/12/11</u> m / d / y	Humidity % :	<u>36.0</u>

Inspection Scale	Actual Readings	
	Clockwise	Counter Clockwise
<u>50</u>	<u>50.20</u>	<u>50.00</u>
<u>150</u>	<u>150.80</u>	<u>150.60</u>
<u>250</u>	<u>250.40</u>	<u>250.60</u>
<u> </u>	<u> </u>	<u> </u>
<u> </u>	<u> </u>	<u> </u>

We certify that the model listed above is calibrated to meet the accuracy and specification requirements as set up by Tohnichi Mfg. Co. Ltd.

Model of Standard Used :	<u>1000DOTE2</u>	Serial No.	<u>XXXXXXW</u>
Last Inspection of Standard :	<u>03/22/11</u>		

Traceable to N.I.S.T. through N.I.S.T. Report No. : 1081768A, B & C
822/266926-02

Inspected at :
Tohnichi America Corporation
677 Academy Drive
Northbrook, IL 60062

Inspected by :
