



Tohnichi America Corp.

1303 Barclay Blvd,

Buffalo Grove, IL 60089

TEL : 847 947 8560

EMAIL : Repairs@Tohnichi.com

CERTIFICATE OF CALIBRATION

Certificate # :	<u>000000-00</u>	Date First Used :	<u> / /</u>
Type of Tool :	<u>TORQUE WRENCH</u>	Accuracy ± :	<u>3%</u>
Model :	<u>QL150F-4A</u>	Serial Number :	<u>XXXXXX</u>
Units :	<u>lbf.ft</u>	Temperature F :	<u>71.0</u>
Date of Calibration : M/D/Y	<u>XXXXXXXX</u>	Humidity % :	<u>53.7</u>

Inspection Scale	Lower	Upper	Actual Readings Result				
<u>30.00</u>	<u>29.2</u>	<u>30.9</u>	CW <u>29.8</u>	<u>29.8</u>	<u>29.8</u>	<u>29.8</u>	<u>29.8</u>
			CCW _____	_____	_____	_____	_____
<u>90.00</u>	<u>87.4</u>	<u>92.7</u>	CW <u>89.4</u>	<u>89.2</u>	<u>89.0</u>	<u>89.0</u>	<u>89.0</u>
			CCW _____	_____	_____	_____	_____
<u>150.00</u>	<u>145.7</u>	<u>154.6</u>	CW <u>148.6</u>	<u>148.8</u>	<u>149.0</u>	<u>148.8</u>	<u>149.0</u>
			CCW _____	_____	_____	_____	_____
_____			_____	_____	_____	_____	_____
			_____	_____	_____	_____	_____

We certify that the model listed above is calibrated to meet the accuracy and specification requirements as set up by Tohnichi Mfg. Co. Ltd.

Model of Standard Used :	<u>TF2000N</u>	Serial No.	<u>706306W</u>
Last Inspection of Standard :	<u>5/3/2017</u>		

Traceable to N.I.S.T. through N.I.S.T. Report No. :	<u>684/286541-15</u>
	<u>684/284451-14</u>

Inspected at :
Tohnichi America Corporation
1303 Barclay Blvd,
Buffalo Grove, IL 60089

Inspected by :
